

CERTIFICATE OF ANALYSIS

PRODUCT	PO5302A	BRILLIANCE " ESB� AGAR
LOT NUMBER	6283840	
EXPIRY DATE	2025.09.09	
TEST DATE	2025.06.17	
REPORTING DATE	2025.06.20	

General Characteristics	Results	Specification
Colour	Conforms	Oyster white
Appearance	Conforms	Opaque
pH	6.9	6.7 -7.1
Packaging / Presentation	Conforms	Conforms to specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Escherichia coli TEM-3 NCTC13351	84	81	1-2mm, blue/ turquoise colonies
Klebsiella pneumoniae SHV-18 ATCC®700603	79	53	1-2mm, green colonies
Enterobacter cloacae ATCC®23355	1.4E+04	0	Complete inhibition (<= 10 cfu)
Citrobacter freundii NCTC8581	1.2E+04	0	Complete inhibition (<= 10 cfu)
Candida albicans ATCC®10231	>100,000	Conforms	No to inhibited growth

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.