

G & M Procter Ltd. Certificate of Analysis

PRODUCT **PO1142A**
MACCONKEY AGAR (EP,USP,JP,BP)
1X10 PLATTEN

LOT NUMBER 4510999
EXPIRY DATE 2025.08.21
PACKING DATE 0000.00.00
TEST DATE 2025.06.18
REPORTING DATE 2025.06.21

All testing in accordance with internally derived specifications, unless otherwise stated.

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Dark red	Dark red	SOP 178 Appearance and colour
pH (25°C)	7.1	6.9 - 7.3	SOP 53 pH
Fill Weight 19 ± 2 Grams	Conforms	Conforms	SOP 74 Fill volume weight check
Contamination @ 32 ± 2°C for >= 72 hours	Conforms	Within acceptable limits	SOP 167 Contamination Check at 22°C & 32°C

MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 50%-150% of the control medium and show the colonial appearance stated in the specification. For partially inhibited organisms, the test medium must achieve the level of inhibition stated in the specification from the stated inoculum.

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Salm. typhimurium ATCC®14028 WDCM-00031	56	59	Straw cols	Straw cols	SOP 151 Fertility of Specified Target Organism(s)(Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



Ian Snowball

Ian Snowball
Quality Manager
G & M Procter Ltd.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,
Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727

G & M Procter Ltd.
Certificate of Analysis

Escherichia coli ATCC®8739	52	51	Dark pink cols	Dark pink cols	SOP 151 Fertility of Specified Target Organism(s)(Agar)
-------------------------------	----	----	----------------	----------------	---

Partially Inhibited Organism(s)	Control (cfu)	Test	Specification	Accredited Method Reference
Enterococcus faecalis ATCC®29212	10,000 - 100,000	Partial to complete	Partial to complete inhibition	SOP 150 Partial Inhibition

EXTRA INFORMATION

All results within this section are reported out with the UKAS ISO 17025 Scope of Accreditation.

Tested in accordance with the harmonized methods described in the current European, United States and Japanese pharmacopoeias for the detection of specified microorganisms in non-sterile products.

*Due to process restriction the packing date may be after the test date.

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



Ian Snowball

Ian Snowball
Quality Manager
G & M Procter Ltd.

CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed

2025.06.21

Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.



Ian Snowball
Quality Manager, G&M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,
4 Auld Bond Road, Perth, PH1 3FX