

G & M Procter Ltd. Certificate of Analysis

PRODUCT PO0520A

TSC SELECTIVE AGAR

1X10 PLATTEN

 LOT NUMBER
 4508542

 EXPIRY DATE
 2025.07.31

 PACKING DATE
 0000.00.00

 TEST DATE
 2025.05.27

 REPORTING DATE
 2025.05.30

All testing in accordance with internally derived specifications, unless otherwise stated.

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Straw 3	Straw 2 or straw 3	SOP 178
			Appearance and colour
pH (25°C)	7.7	7.4 - 7.8	SOP 53 pH
Fill Volume/Weight	17.9g	17.0 - 19.0g	SOP 74 Fill volume weight check
Contamination @ 32 ± 2 °C for >= 72 hours	Conforms	Within acceptable limits	SOP 167 Contamination Check at 22°C & 32°C

MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 70%-150% of the control medium and show the colonial appearance stated in the specification.

Test Organism	Control c.f.u.	Test Result	Specification	Accredited Method Reference
Clostridium perfringens ATCC®13124	44	46	Black cols with halo	SOP 151 Fertility of Specified Target Organism(s)(Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.





gan Snorbell

Ian Snowball Quality Manager G & M Procter Ltd.



G & M Procter Ltd. Certificate of Analysis

Inhibited Organism	Control(cfu)	Test	Specification	Accredited Method Reference
Escherichia coli ATCC®25922	10,000 - 100,000	No growth	No growth	SOP 155 Inhibition

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.





gan Snorbell

Ian Snowball
Quality Manager
G & M Procter Ltd.