

## Certificate of Analysis

**Product Name:** PathoDx Strep Grouping Kit KT/60TST  
**Product Code:** R62025  
**Lot Number:** 6206518  
**Expiration Date:** 23.09.2026  
**Date of Manufacture:** 02.06.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75PDX04E	Strep A Group latex Filled bulk	6193759	23.09.2026
W74PDX13N	Extraction Reagent 1 Filled Bulk	6193760	19.03.2027
W74PDX15R	7ml Extraction Reagent 2 Filled Bulk	6193763	20.03.2027
W74PDX17T	14ml Extraction Rgt 3 Filled Bulk	6193761	04.02.2027
W75PDX06G	Strep B Group Latex Filled Bulk	6209743	05.10.2026
W75PDX07H	Strep C Group Latex Filled Bulk	6209744	29.09.2026
W75PDX09K	Strep F Group Latex Filled Bulk	6209745	18.10.2026
W75PDX10L	Strep G Group Latex Filled Bulk	6209746	09.12.2026
W75PDX01B	PathoDx Pos. Control Filled Bulk	6193764	05.02.2027

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Compatibility of Matched Kit Components	Each Latex must show obvious agglutination with its homologous extract	Conforms
	Each Latex must not show agglutination with the heterologous extracts although a trace reaction is acceptable	Conforms
	Each Latex must show strong positive agglutination with its homologous positive control antigen extract	Conforms
	Each Latex must not show agglutination with its heterologous positive control antigen extract although a trace reaction is acceptable	Conforms



Fay Curd  
Quality Assurance Manager