

Certificate of Analysis

Product Name: PathoDx Strep B Latex KT/60TST
Product Code: R62031
Lot Number: 6236735
Expiration Date: 05.10.2026
Date of Manufacture: 02.05.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75PDX06G	Strep B Group Latex Filled Bulk	6209743	05.10.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Latex Efficacy/ Specificity using Extraction Reagents	The Latex must show obvious agglutination with its homologous extract	Conforms
	The Latex must not show agglutination with the heterologous extracts but a trace reaction is acceptable	Conforms
Reactivity of Latex with Control Antigens ABC and FG	The Latex must show strong positive agglutination with its homologous positive control antigen extract	Conforms
	The Latex must show no agglutination with its heterologous positive control antigen extract	Conforms



Fay Curd
 Quality Assurance Manager