

## Certificate of Analysis

**Product Name:** BactiStaph Latex 150TST/KT  
**Product Code:** R21144  
**Lot Number:** 6223262  
**Expiration Date:** 07.07.2026  
**Date of Manufacture:** 03.06.2025

\*Date format is DD.MM.YYYY

| Item      | Description                             | Lot/Serial | Expire     |
|-----------|-----------------------------------------|------------|------------|
| W77BCL40S | BactiStaph Latex Reagent Filled Bulk    | 6222999    | 16.07.2026 |
| W77BCL41T | BactiStaph Positive Control Filled Bulk | 6223000    | 13.07.2026 |
| W77BCL42U | BactiStaph Negative Control Filled Bulk | 6223261    | 07.07.2026 |

### Organism Challenge

Staphylococcus aureus (3 strains\*) (+) Agglutination  
 Coagulase (-) Staphylococcus (3 strains)  
 Positive Control  
 Negative Control

\* 1 = MRSA

### Physical Characteristics:

Appearance:  
 pH @ 20-25°C:

### Performance Criteria

(-) Agglutination  
 (+) Agglutination  
 (-) Agglutination

Black Liquid  
 N/A

Microbial Load Evaluation:  
 N/A



Fay Curd  
 Quality Assurance Manager