

Certificate of Analysis

Product Name: Salmonella H Poly Phases 1&2-2ml/VL
Product Code: R30858501
Lot Number: 6247689
Expiration Date: 31.05.2028
Date of Manufacture: 06.06.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75BZD01E	Salmonella Poly H 1-2 Filled Bulk	6222063	31.05.2028

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Anticipated result	Actual Result
Abbreviated Slide Agglutination Test	Saline control must not show agglutination	Conforms
	Homologous controls must show agglutination equal to or greater than +	Conforms
	Serum under test must show agglutination equal to or greater than + with the homologous culture	Conforms
	Serum under test must not show agglutination with the heterologous culture	Conforms



Fay Curd
Quality Assurance Manager