

CERTIFICATE OF ANALYSIS

PRODUCT SR0059C
MYCOPLASMA SUPPLEMENT G 1 X 10 VIALS

LOT NUMBER 6242984

EXPIRY DATE 2027.03.26

**DATE OF
MANUFACTURE** 2025.03.26

Delivery/Customer information

Date Printed

2025.05.01

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Brown pellet	Brown pellet

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	----------------	-------------	--------------------	-------------

Tested in Mycoplasma Agar Base CM0401

Microaerophilic incubation at 37°C for up to 21 days

<i>Mycoplasma hominis</i> ATCC®23114	>1E+04	-	-	Characteristic 'fried egg' cols
<i>Mycoplasma pneumoniae</i> ATCC®15531	>1E+04	-	-	Characteristic 'fried egg' cols
<i>Mycoplasma bovis</i> ATCC®19852	>1E+04	-	-	Characteristic 'fried egg' cols
<i>Mycoplasma orale</i> ATCC®23714	>1E+04	-	-	Characteristic 'fried egg' cols
<i>Acholeplasma laidlawii</i> ATCC®23206	>1E+04	-	-	Characteristic 'fried egg' cols

Microaerophilic incubation at 37°C for 5 days

<i>Streptococcus pyogenes</i> ATCC®19615	>1E+04	-	-	No growth
<i>Staphylococcus aureus</i> ATCC®25923	>1E+04	-	-	No growth
<i>Proteus mirabilis</i> ATCC®12453	>1E+04	-	-	No growth
	Control cfu	Test cfu	Log(10) Reduction	
<i>Escherichia coli</i> ATCC®25922	5.3E+04	0.0E+00	5	No growth

Control Medium: Mycoplasma Agar Base with the addition of Mycoplasma Supplement G or Tryptone Soya Agar, where appropriate

A satisfactory result is represented by characteristic colonies from an inoculum of 1E+04 to 1E+06 colony-forming units (cfu) after microscopic examination using x40 magnification.

Refer to product specification for full details.

CERTIFICATE OF ANALYSIS

PRODUCT	SR0059C
	MYCOPLASMA SUPPLEMENT G 1 X 10 VIALS
LOT NUMBER	6242984

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.01

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.