

CERTIFICATE OF ANALYSIS

PRODUCT TV5065E MKTTN ENRICHMENT BROTH (50X10ML)
LOT NUMBER 6272856
EXPIRY DATE 2025.08.18
TEST DATE 2025.05.27
REPORTING DATE 2025.06.02

| General Characteristics | Results | Specification |
|--|----------|---------------------|
| Colour | Conforms | Light green |
| Appearance | Conforms | Opaque |
| End of manufacturing pH | 7.9 | 7.8 - 8.2 |
| Packaging / Presentation | Conforms | Label & Print check |
| Filling volume / weight | Conforms | 9,0 - 11,0 g |
| Cont. check @ 20-25 & 30-35°C for >= 72h | Conforms | Within Limits |

| Microbiological Performance | Control c.f.u | Test Result | Specification |
|--|---------------|-------------|---------------------------------|
| Mixed culture of Salmonella Typhimurium ATCC®14028 WDCM 00031 and 10.000 - 100.000 cfu / tube | 75 | Conforms | >10 colonies on XLD Agar |
| Pseudomonas aeruginosa ATCC®27853 WDCM 00025 | 1E+04 - 1E+05 | Conforms | No growth on XLD Agar |
| Escherichia coli ATCC®25922 WDCM 00013 | 1E+04 - 1E+05 | Conforms | No growth on XLD Agar |
| Pure cultures Escherichia coli ATCC®25922 WDCM 00013 | 1E+04 - 1E+05 | Conforms | Partial inhibition (<= 100 cfu) |
| Enterococcus faecalis ATCC®29212 WDCM 00087 | 1E+04 - 1E+05 | Conforms | Complete inhibition (<= 10 cfu) |

Biological performance tested in accordance with current version of ISO 11133.

The formulation conforms to ISO 6579.

Control Media: XLD Agar and Tryptone Soya Agar

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.

CERTIFICATE OF ANALYSIS

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkkS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.