

## CERTIFICATE OF ANALYSIS

**PRODUCT** TV5065E MKTTN ENRICHMENT BROTH (50X10ML)  
**LOT NUMBER** 6276440  
**EXPIRY DATE** 2025.08.25  
**TEST DATE** 2025.06.02  
**REPORTING DATE** 2025.06.06

General Characteristics	Results	Specification
Colour	Conforms	Light green
Appearance	Conforms	Opaque
End of manufacturing pH	7.8	7.8 - 8.2
Packaging / Presentation	Conforms	Label & Print check
Filling volume / weight	Conforms	9,0 - 11,0 g
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Mixed culture of Salmonella Typhimurium ATCC®14028 WDCM 00031 and 10.000 - 100.000 cfu / tube	67	Conforms	>10 colonies on XLD Agar
Pseudomonas aeruginosa ATCC®27853 WDCM 00025	1E+04 - 1E+05	Conforms	No growth on XLD Agar
Escherichia coli ATCC®25922 WDCM 00013	1E+04 - 1E+05	Conforms	No growth on XLD Agar
Pure cultures Escherichia coli ATCC®25922 WDCM 00013	1E+04 - 1E+05	Conforms	Partial inhibition (<= 100 cfu)
Enterococcus faecalis ATCC®29212 WDCM 00087	1E+04 - 1E+05	Conforms	Complete inhibition (<= 10 cfu)

Biological performance tested in accordance with current version of ISO 11133.

The formulation conforms to ISO 6579.

Control Media: XLD Agar and Tryptone Soya Agar

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

The quality control methods meet requirements of ISO 11133.

## CERTIFICATE OF ANALYSIS

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkkS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.