

Certificate of Analysis

Product Name: PathoDxtra Strep Grouping Reagent Set
Product Code: DR0710M
Lot Number: 6215337
Expiration Date: 31.10.2026
Date of Manufacture: 06.06.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74PDX76H	Pathodxtra Positive Control	6194087	28.02.2027
W74PDX77J	Reagent 1	6186663	31.10.2026
W74PDX78K	Reagent 2	6186664	31.10.2026
W74PDX79L	Reagent 3	6186665	31.10.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Efficacy of Extraction Reagents 1, 2 and 3	Each Latex must show obvious agglutination with its homologous extracts	Conforms
	Each Latex must not show agglutination with the heterologous extracts, nor with the extraction fluid from the uninoculated tube although trace granularity is acceptable	Conforms
Reactivity of Latexes with Polyvalent Positive Control	The Polyvalent Positive Control must show obvious agglutination with each Latex and the reactions must be of roughly equivalent strength	Conforms



Fay Curd
Quality Assurance Manager