

Certificate of Analysis

Product Name: PathoDextra Extraction Reagents
Product Code: DR0709M
Lot Number: 6275339
Expiration Date: 31.01.2027
Date of Manufacture: 22.09.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74PDX77J	Reagent 1	6224752	31.01.2027
W74PDX78K	Reagent 2	6224751	31.01.2027
W74PDX79L	Reagent 3	6224749	31.01.2027

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Efficacy of Extraction Reagents 1, 2 and 3	Each Latex must show obvious agglutination with its homologous extracts	Conforms
	Each Latex must not show agglutination with the heterologous extracts, nor with the extraction fluid from the uninoculated tube although trace granularity is acceptable	Conforms



Fay Curd
Quality Assurance Manager