

## Certificate of Analysis

**Product Name:** PathoDextra Strep Group G Latex  
**Product Code:** DR0706G  
**Lot Number:** 6197389  
**Expiration Date:** 30.09.2026  
**Date of Manufacture:** 29.05.2025

\*Date format is DD.MM.YYYY

| Item      | Description         | Lot/Serial | Expire     |
|-----------|---------------------|------------|------------|
| W74PDX75G | Strep G Group Latex | 6194088    | 30.09.2026 |

Test results carried out according to the Quality Assurance Product Release and Test Specification

| Test                                                        | Anticipated Result                                                                                      | Actual Result |
|-------------------------------------------------------------|---------------------------------------------------------------------------------------------------------|---------------|
| Latex Efficacy/<br>Specificity using<br>Extraction Reagents | The Latex must show obvious agglutination with its homologous extract                                   | Conforms      |
|                                                             | The Latex must not show agglutination with the heterologous extracts but a trace reaction is acceptable | Conforms      |
| Reactivity of Latex with Positive Control                   | The Latex must show strong positive agglutination with the positive control.                            | Conforms      |



Fay Curd  
Quality Assurance Manager