

CERTIFICATE OF ANALYSIS

PRODUCT PO0678D PRO-TECT C/PLATE (X3 WRAP + IRRAD)

LOT NUMBER 6281056

EXPIRY DATE 2025.09.16 **TEST DATE** 2025.06.16 **REPORTING DATE** 2025.06.23

General Characteristics	Results		Specification
Colour	Conforms		Straw
Appearance	Conforms		Transparent
pH	7.2		7.1 -7.5
Packaging / Presentation	Conforms		10 plates, triple-wrapped
Average irradiation dose (kGy)	14.83		11.50 - 22.00
Cont. check @ 20-25 & 30-35°C for >=120h	Conforms		No Contamination
Microbiological Performance	Control c.f.u	Test Result	Specification
Incubation @ 30-35°C up to 72h, aerobic			
Escherichia coli ATCC®8739	73	85	3-5mm, cream colonies
Staphylococcus aureus ATCC®6538	82	87	1-3mm, straw colonies
Pseudomonas aeruginosa ATCC®9027	44	38	2-5mm, straw colonies
Bacillus subtilis ATCC®6633	70	82	3-9mm, irregular straw colonies
Incubation @ 32°C up to 48h, anaerobic			
Clostridium sporogenes ATCC®19404	56	60	1-2 mm, grey to cream coloniess
Incubation @ 20-25°C up to 72h, aerobic			•
Bacillus subtilis ATCC®6633	70	79	3-9mm, irregular straw colonies
Incubation @ 20-25°C up to 120h, aerobic			
Candida albicans ATCC®10231	63	70	2-3mm, cream colonies
Aspergillus brasiliensis ATCC®16404	25	30	10-30mm, white mycelia, black spores

Colony counts shall be equal to or greater than 50% of the control medium (Tryptone Soya Agar or Sabouraud Dextrose Agar)

Tested in accordance with the methods described in the current United States pharmacopoeia for the microbiological control and monitoring of aseptic processing environments.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

DAKKS

Deutsche
Akkreditrungsstelle
DP-120199-01-00

For microbiology