

Certificate of Analysis - Accredited Reference Material

Thermo Scientific TM	Frademark TM
Product Number	R4715210
Product Name	P. paraeruginosa ATCC 9027 Q+ KT/100 TST
Lot Number	26 9 876
Usage Decision	Accepted (OK)
Expiration Date	2026-05-22

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. The results were derived from a representative sample of the batch and were obtained at the time of release. Refer to the enclosed product insert for instructions, intended use, hazard/safety requirements, and storage conditions.

Product Char Results

Purity	Demonstrates pure growth on applicable media
Viability	Recovered at acceptable level within test period
Passage	2 (Current preserved state)
Delivery	<100 CFU/0.1 mL

Microbiological Testing	Results	Specification
>85% Identification on Vitek 2C GN >95% Identification on MicroSEQ	98	85 - 100 95 - 100
CFU Mean	75	-
Microscopic Features	Pass	

These tests are performed in accordance with ISO 17025 guidelines.

Thermo Fisher Scientific has determined each loop of this reference material to be sufficiently homogeneous for its intended use.

Individual products are traceable to a recognized culture collection.

Although the Vitek(TM) panel uses many conventional tests, the unique environment of the card, combined with the short incubation period, may produce results that differ from published results obtained by other methods

Signed

Abh

Dawn Baker QA Manager

The identity, purity, and authenticity of the Licensed Products are exclusively the responsibility of Remel Inc. and not ATCC. The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative Word mark, and the ATCC Catalog Marks are trademarks of ATCC.

Remel Inc. is licensed to use these trademarks and to sell products derived from ATCC® culture.





Thermo Fisher Scientific is accredited by A2LA as a registered reference material producer certificate number 6559.01 in accordance with ISO 17034[1].

[1] ISO 17034 First Edition 2016-11-01 General requirements for the competence of reference material producers

Version 1.0

Thermo Fisher Scientific Laboratory Report

Isolate: 269876-1 (To be reviewed) Card Type: GN Bar Code: 2412977103335708 Testing Instrument: 00000B4D4CA6 (01783) Setup Technologist: Amanda Wood(awood)

Bionumber: 0003451303500242 Organism Quantity:

Selected Organism: Pseudomonas aeruginosa

C	
Comments:	

McFarland:									
Identification	Card: GN		Lot Number:	2412977103	Expires:	Jan 20, 2026 12:00 CST			
Information	Status:	Final	Analysis Time:	5.85 hours	Completed:	May 29, 2025 21:04 CDT			
Organism Origin	VITEK 2								
Selected Organism	98% Prob Bionumb	oability er: 00034513	Pseudomona 03500242	8	lence: Excellent	identification			
Analysis Organisms and Tests to Separate:									
Analysis Messages:									

Contraindicating Typical Biopattern(s)

Pseudomonas aeruginosa URE(16),

Bioo	chemical D	etails															
2	APPA	-	3	ADO	-	4	PyrA	-	5	lARL	-	7	dCEL	-	9	BGAL	-
10	H2S	-	11	BNAG	-	12	AGLTp	-	13	dGLU	+	14	GGT	+	15	OFF	-
17	BGLU	-	18	dMAL	-	19	dMAN	+	20	dMNE	+	21	BXYL	-	22	BAlap	+
23	ProA	+	26	LIP	-	27	PLE	-	29	TyrA	+	31	URE	+	32	dSOR	-
33	SAC	-	34	dTAG	-	35	dTRE	-	36	CIT	+	37	MNT	+	39	5KG	-
40	lLATk	+	41	AGLU	-	42	SUCT	+	43	NAGA	-	44	AGAL	(-)	45	PHOS	-
46	GlyA	-	47	ODC	-	48	LDC	-	53	lHISa	-	56	CMT	+	57	BGUR	-
58	O129R	-	59	GGAA	-	61	lMLTa	+	62	ELLM	-	64	lLATa	+			

bioMérieux Customer: 1011208 System #: 504538 Thermo Fisher Scientific Laboratory Report

Autoprint Report Version: 1 of 1

Isolate: 269876-1 (To be reviewed) Card Type: GN Bar Code: 2412977103335708 Testing Instrument: 00000B4D4CA6 (01783) Setup Technologist: Amanda Wood(awood)

Bionumber: 0003451303500242 Organism Quantity:

Selected Organism: Pseudomonas aeruginosa

Action

Name (User ID)

Date/Time

Comment

Installed VITEK 2 Systems Version: 9.02 MIC Interpretation Guideline: AES Parameter Set Name:

Therapeutic Interpretation Guideline: AES Parameter Last Modified: Page 2 of 2