



Certificate of Quality

Product Name: M4RT(3ML) PK/72
Lot Number: 258974

Product Number: R12505
Expiration Date: 2027-04-10
(YYYY-MM-DD)


This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	Strain	Results
Maintenance:		
Chlamydia trachomatis	VR-880	Pass
Herpes simplex type 2	VR-734	Pass
Cytomegalovirus	VR-977	Pass
Respiratory syncytial virus	VR-1540	Pass

Toxicity:	
MRC5 Cells	CCL-171 Pass
VERO Cells	CCL-81 Pass
McCoy Cells	CRL-1696 Pass

Physical Characteristics:	
Appearance:	Orange-Pink Clear Broth with Clear Beads
pH @ 20-25°C:	7.3 +/- 0.2

Microbial Load Evaluation:
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed


Senior Quality Assurance Engineer