

CERTIFICATE OF ANALYSIS

PRODUCT	PO5170A	BRILLIANCE (TM) CANDIDA
LOT NUMBER	6270881	
EXPIRY DATE	2025.07.24	
TEST DATE	2025.05.15	
REPORTING DATE	2025.05.20	

General Characteristics	Results	Specification
Colour	Conforms	Oyster white
Appearance	Conforms	Opaque
pH	6.0	5.8 -6.2
Packaging / Presentation	Conforms	Conforms to specification
Cont. check @ 20-25 & 30-35°C for >=120h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Candida albicans ATCC®10231	52	56	1 - 2mm, green colonies
Escherichia coli ATCC®25922	1.7E+04	0	Partial inhibition (<= 100 cfu)
Candida tropicalis ATCC®750	<10000	Conforms	Good growth, dark blue colonies
Candida krusei ATCC®6258	<10000	Conforms	Good growth, dry irregular pink brown col

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.