

CERTIFICATE OF ANALYSIS

PRODUCT	PB5253E	COLUMBIA AGAR+S-BLOOD/BRILLIANCE MRSA 2
LOT NUMBER	6289590	
EXPIRY DATE	2025.08.28	
TEST DATE	2025.07.04	
REPORTING DATE	2025.07.08	

General Characteristics Side 1	Results	Specification
Colour	Conforms	Traffic red
Appearance	Conforms	Opaque
pH	7.4	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check

Microbiological Performance Side 1	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC®33591	<10000	Conforms	Good growth, white cols with haemolysis
Streptococcus pneumoniae ATCC®6305	<10000	Conforms	Good growth, grey cols, alpha haemolysis
Streptococcus pyogenes ATCC®12344	<10000	Conforms	Good growth, grey cols, beta haemolysis
Pseudomonas aeruginosa ATCC®27853	<10000	Conforms	Good growth, grey shiny colonies

General Characteristics Side 2	Results	Specification
Colour	Conforms	Light ivory
Appearance	Conforms	Opaque
pH	7.4	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check

Microbiological Performance Side 2	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC®33591	<10000	Conforms	Good growth, blue colonies
Staphylococcus aureus ATCC®29213	>100,000	Conforms	No growth
Proteus mirabilis ATCC®29906	>100,000	Conforms	No growth
Pseudomonas aeruginosa ATCC®27853	>100,000	Conforms	No growth
Bacillus licheniformis ATCC®14580	<10000	Conforms	Growth of small, rose colonies

3 Day - Contamination Check	Results	Specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.

CERTIFICATE OF ANALYSIS

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkks according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.