

Certificate of Quality

Product Name: SUPPLEMENTED BRUCELLA BROTH
Lot Number: 256532


Product Number: T3450
Expiration Date: 2025-12-29
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Bacteroides thetaiotaomicron	29741	MIC limits are within CLSI specifications
Bacteroides fragilis	25285	MIC limits are within CLSI specifications

Physical Characteristics:
Appearance: Red, Clear Broth w/ No Significant Precipitate
pH @ 20-25°C: 7.3+/- 0.05

Microbial Load Evaluation:
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 7 days at 20-25C and 33-37C.

Signed


Senior Quality Assurance Engineer