

CERTIFICATE OF ANALYSIS

PRODUCT	PO5172C	TSA WITH DISINHIBITOR PLUS
LOT NUMBER	6281092	
EXPIRY DATE	2025.09.04	
TEST DATE	2025.06.13	
REPORTING DATE	2025.06.18	

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	7.4	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >=120h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Incubation @ 30-35°C up to 72h, aerobic			
Escherichia coli ATCC®8739	76	68	3-5mm, transparent colonies
Staphylococcus aureus ATCC®6538	57	55	1-3mm, orange shiny colonies
Pseudomonas aeruginosa ATCC®9027	89	105	2-5mm, cream colonies
Bacillus subtilis ATCC®6633	57	65	3-9mm, cream colonies
Incubation @ 20-25°C up to 72h, aerobic			
Bacillus subtilis ATCC®6633	57	61	3-9mm, cream colonies
Incubation @ 20-25°C up to 120h, aerobic			
Candida albicans ATCC®10231	75	81	2-3mm, cream colonies
Aspergillus brasiliensis ATCC®16404	19	21	10-30mm white mycelium, black spores

Colony counts shall be equal to or greater than 50% of the control medium (Tryptone Soya Agar or Sabouraud Dextrose Agar)

Growth promotions tests are derived from the United States pharmacopoeia for the microbiological control and monitoring of aseptic processing environments.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.