

## G & M Procter Ltd. Certificate of Analysis

**PRODUCT**                      **BO0369M**  
**TRYPTONE SOYA BROTH (acc. to EP/USP)**  
**10X100ML IN 100ML VIAL**

**LOT NUMBER**                4508343  
**EXPIRY DATE**                2026.05.26  
**PACKING DATE**              2025.05.22  
**TEST DATE**                    2025.05.23  
**REPORTING DATE**            2025.06.06

All testing in accordance with internally derived specifications, unless otherwise stated.

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Straw 2-3	Straw 2, straw 2-3 or straw 3	SOP 183 Appearance and colour
pH (25°C)	7.5	7.1 - 7.5	SOP 53 pH
Fill Volume/Weight	100.6g	100.0 - 101.5g	SOP 154 Fill volume weight check
Sterility @ 22°C & 32°C ± 2°C for 14days	no growth	no growth	SOP 167 Contamination Check at 22°C & 32°C

### MICROBIOLOGICAL PERFORMANCE

For target organisms, the control media must achieve a colony count of 10-100 cfu. The test medium must show clearly visible growth as stated in the specification.

Target Organism	Control(cfu)	Test	Specification	Accredited Method Reference
Staphylococcus aureus ATCC®6538	84	Turbid growth	Turbid growth	SOP 151 Fertility of Specified Target Organism(s) (Liquid Media)
Pseudomonas aeruginosa ATCC®9027	55	Turbid growth	Turbid growth	SOP 151 Fertility of Specified Target Organism(s) (Liquid Media)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



*Ian Snowball*

Ian Snowball  
Quality Manager  
G & M Procter Ltd.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,  
Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727

## G & M Procter Ltd. Certificate of Analysis

Escherichia coli ATCC®8739	81	Turbid growth	Turbid growth	SOP 151 Fertility of Specified Target Organism(s) (Liquid Media)
Bacillus subtilis ATCC®6633	94	Flocculent/surface growth	Flocculent/surface growth	SOP 151 Fertility of Specified Target Organism(s) (Liquid Media)

Target Organism	Control(cfu)	Test	Specification	Accredited Method Reference
Candida albicans ATCC®10231	88	Flocculent/surface growth	Flocculent/surface growth	SOP 151 Fertility of Specified Target Organism(s) (Liquid Media)
Aspergillus brasiliensis ATCC®16404	83	White mycelia, black spores	White mycelia with/without black spores	SOP 151 Fertility of Specified Target Organism(s) (Liquid Media)
32°C B.subtilis ATCC®6633	96	Turbid growth	Turbid growth	SOP 151 Fertility of Specified Target Organism(s) (Liquid Media)
Salm. typhimurium ATCC®14028 WDCM-00031	88	Turbid growth	Turbid growth	SOP 151 Fertility of Specified Target Organism(s) (Liquid Media)

### EXTRA INFORMATION

#### All results within this section are reported out with the UKAS ISO 17025 Scope of Accreditation.

The product was tested in accordance with the current versions of the European Pharmacopoeia, the British Pharmacopoeia, the US Pharmacopoeia and the Japanese Pharmacopoeia. Clearly visible growth was observed within 3 days for bacteria and within 5 days for fungi

\*Due to process restrictions the packing date may be after the test date

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



*Ian Snowball*

Ian Snowball  
Quality Manager  
G & M Procter Ltd.

## CERTIFICATE OF ANALYSIS

### Delivery/Customer information

Date Printed

2025.06.06

Delivery No.

Customer

Customer Order number

---

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.



Ian Snowball  
Quality Manager, G&M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,  
4 Auld Bond Road, Perth, PH1 3FX