

CERTIFICATE OF ANALYSIS

PRODUCT CM0329B
BRILLIANT GREEN AGAR (Modified) 500g

LOT NUMBER 6257159

EXPIRY DATE 2030.05.31

DATE OF MANUFACTURE 2025.05.12

Delivery/Customer information

Date Printed

2025.06.17

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Light green or orange powder	Light green or orange powder
Colour on reconstitution	Orange/red	Orange/red
pH (25°C)	6.8	6.7 - 7.1
Clarity	Clear	Clear

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
Aerobic incubation at 37°C for 18 hours				
<i>Salmonella enteritidis</i> ATCC®13076	85	81	95	Red cols & medium
<i>Salmonella typhimurium</i> ATCC®14028	66	52	79	Red cols & medium
<i>Salmonella virchow</i> NCTC5742	77	59	77	Red cols & medium
<i>Salmonella nottingham</i> NCTC7832	48	59	123	Red cols & medium
<i>Pseudomonas aeruginosa</i> ATCC®9027	82	87	106	Red cols & medium
<i>Escherichia coli</i> ATCC®25922	>1E+04	-	-	Yellow/green cols
Log(10) Reduction				
<i>Escherichia coli</i> ATCC®11775	6.9E+04	0.0E+00	5	No growth
<i>Proteus mirabilis</i> ATCC®29906	6.9E+04	3.0E-01	5	Pink cols, no swarming
<i>Proteus mirabilis</i> ATCC®12453	7.7E+04	0.0E+00	5	No growth
<i>Enterobacter cloacae</i> ATCC®13047	4.7E+04	5.8E+01	3	Yellow/green cols

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu). Negative strains are inhibited or shall produce at least a 1 log(10) reduction when compared to the control medium.

For *Pseudomonas aeruginosa* ATCC®9027, a satisfactory result is represented by recovery equal to or greater than 50% of the control medium from an inoculum of 10-100 cfu. For *Escherichia coli* ATCC®25922, a satisfactory result is represented by a negative diagnostic reaction from an inoculum of 1E+04 to 1E+06 cfu.

Refer to product specification for full details.

CERTIFICATE OF ANALYSIS

PRODUCT	CM0329B
LOT NUMBER	BRILLIANT GREEN AGAR (Modified) 500g 6257159

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.09

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.