

CERTIFICATE OF ANALYSIS

PRODUCT CM0331T
COLUMBIA BLOOD AGAR BASE 5kg

LOT NUMBER 6323600

EXPIRY DATE 2030.10.31

DATE OF MANUFACTURE 2025.09.27

Delivery/Customer information

Date Printed

2025.10.17

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 2	Straw 2-3
pH (25°C)	7.2	7.1 - 7.5
Clarity	Clear	Clear
Blood compatibility with horse blood	Satisfactory	Satisfactory
Blood compatibility with sheep blood	Satisfactory	Satisfactory

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
Aerobic incubation at 37°C for 24 hours				
<i>Staphylococcus aureus</i> ATCC®25923	50	56	112	Cream cols
Anaerobic incubation at 37°C for 48 hours				
<i>Clostridium sporogenes</i> ATCC®19404	50	54	108	Pale straw cols
<i>Clostridium sporogenes</i> ATCC®11437	32	24	75	Pale straw cols
Enriched with 5% v/v horse blood				
Aerobic incubation at 37°C for 24 hours				
<i>Streptococcus pyogenes</i> ATCC®19615	34	34	102	Pale straw cols, beta haemolysis
<i>Streptococcus pneumoniae</i> ATCC®6305	20	20	100	Grey/green cols, alpha haemolysis
Enriched with 5% v/v horse blood				
Microaerophilic incubation at 37°C for 48 hours				
<i>Neisseria gonorrhoeae</i> NCTC11148	93	94	101	Grey/brown cols
Growth of <i>Haemophilus</i> spp. in the presence of X & V factor discs				
Aerobic incubation at 37°C for 18 hours				
	X Factor (mm)	V Factor (mm)	X+V Factor (mm)	
<i>Haemophilus influenzae</i> ATCC®49247	0	0	16	
<i>Haemophilus parainfluenzae</i> ATCC®33392	0	47	42	

CERTIFICATE OF ANALYSIS

PRODUCT CM0331T
COLUMBIA BLOOD AGAR BASE 5kg
LOT NUMBER 6323600

Enriched with 7% v/v horse blood
Bacitracin activity
Aerobic incubation at 37°C for 18 hours

	Zone Size (mm)	Limits (mm)
<i>Streptococcus pyogenes</i> ATCC®19615	18	10 - 20

Testing has been performed in accordance with ISO11133:2014

Enriched with 5% v/v sheep blood

Microaerophilic incubation at 41.5 ± 2°C for 44 ± 4 hours

<i>C. jejuni</i> ATCC®29428 WDCM00156	81	93	115	Grey, mucoid cols
<i>C. jejuni</i> ATCC®33291 WDCM00005	90	92	102	Grey, mucoid cols
<i>C. coli</i> ATCC®43478 WDCM00004	70	72	103	Grey, mucoid cols
	Initial Inoculum		Test Result	
<i>Streptococcus pyogenes</i> ATCC®19615	220			Pale straw cols, beta haemolysis
<i>Streptococcus pneumoniae</i> ATCC®6305	320			Grey/green cols, alpha haemolysis
<i>Staphylococcus aureus</i> ATCC®25923	600			Cream cols
<i>Escherichia coli</i> ATCC®25922	370			Cream cols

Control Medium: Columbia Blood Agar Base

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.
Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.10.17

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.