

CERTIFICATE OF ANALYSIS

PRODUCT CM0929B
MYP AGAR 500g

LOT NUMBER 6259346

EXPIRY DATE 2030.04.30

DATE OF MANUFACTURE 2025.04.14

Delivery/Customer information

Date Printed

2025.04.28

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Red	Red
pH (25°C)	7.3	7.0 - 7.4
Clarity	Clear	Clear

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	----------------	-------------	--------------------	-------------

Tested with the addition of 10% v/v Egg Yolk Emulsion SR0047 and Polymyxin B Supplement SR0099

Aerobic incubation at 30 ± 2°C for 24 ± 3 hours				
<i>Bacillus cereus</i> ATCC®10876	58	42	73	Bright pink cols with halo

Aerobic incubation at 30 ± 2°C for 44 ± 4 hours				
<i>Pseudomonas aeruginosa</i> ATCC®27853	>1E+04	0	-	No growth

Testing has been performed in accordance with ISO11133:2014

Aerobic incubation at 30 ± 2°C for 24 ± 3 hours				
<i>B. cereus</i> ATCC®11778 WDCM00001	54	47	87	Bright pink cols with halo

Aerobic incubation at 30 ± 2°C for 44 ± 4 hours				
<i>E. coli</i> ATCC®25922 WDCM00013	>1E+04	0	-	No growth
<i>E. coli</i> ATCC®8739 WDCM00012	>1E+04	0	-	No growth
	Inoculum			Test Description
<i>B. subtilis</i> ATCC®6633 WDCM00003	5.2E+03	-	-	Yellow/orange cols, no halo

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.
Refer to product specification for full details.

CERTIFICATE OF ANALYSIS

PRODUCT	CM0929B	
	MYP AGAR	500g
LOT NUMBER	6259346	

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.04.28

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com