

CERTIFICATE OF ANALYSIS

PRODUCT CM1135B
BRAIN HEART INFUSION BROTH 500g

LOT NUMBER 6246645

EXPIRY DATE 2030.04.30

DATE OF MANUFACTURE 2025.03.20

Delivery/Customer information

Date Printed

2025.04.28

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 3-4	Straw 3-4
pH (25°C)	7.5	7.2 - 7.6
Clarity	Clear	Clear

Microbiological Performance	Initial Inoculum	Test Result
Aerobic incubation at 37 ± 2°C for 18 hours		
<i>Streptococcus pyogenes</i> ATCC®19615	58	Turbid growth
<i>Streptococcus pneumoniae</i> ATCC®6303	39	Turbid growth
<i>Streptococcus pneumoniae</i> ATCC®6305	25	Turbid growth
<i>Enterococcus faecalis</i> ATCC®19433	25	Turbid growth
<i>Pseudomonas aeruginosa</i> ATCC®27853	48	Turbid growth
Anaerobic incubation at 37 ± 2°C for 18 hours		
<i>Streptococcus pneumoniae</i> ATCC®6305	25	Turbid growth
Aerobic incubation at 37 ± 2°C for 48 hours		
<i>Candida albicans</i> ATCC®10231	46	Turbid growth
Addition of 0.1% w/v bacteriological agar		
Aerobic incubation at 37 ± 2°C for 48-72 hours		
<i>Bacteroides fragilis</i> ATCC®25285	38	Turbid growth
Enriched with X & V factor discs		
Aerobic incubation at 37 ± 2°C for 24 hours		
<i>Haemophilus influenzae</i> ATCC®33391	24	Turbid growth
Enriched with 10% v/v horse blood		
Aerobic incubation at 37 ± 2°C for 48 hours and subculture onto chocolate agar		
<i>Neisseria meningitidis</i> ATCC®13077	67	Grey/brown cols
<i>Neisseria gonorrhoeae</i> NCTC11148	66	Grey/brown cols
<i>Neisseria gonorrhoeae</i> ATCC®19424	24	Grey/brown cols

Tube coagulase test

Aerobic incubation at 37 ± 2°C for 4 hours

Staphylococcus aureus ATCC®9144 1E+04-1E+06 Positive

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com

CERTIFICATE OF ANALYSIS

PRODUCT CM1135B
BRAIN HEART INFUSION BROTH 500g
LOT NUMBER 6246645

Staphylococcus epidermidis ATCC®14990 1E+04-1E+06 Negative

Testing has been performed in accordance with ISO11133:2014

Aerobic incubation at 37 ± 2°C for 24 ± 2 hours

Staphylococcus aureus ATCC®25923 78 Turbid growth

Tested in accordance with current CLSI M22 A

Aerobic incubation at 33-37°C for 18 hours

Escherichia coli ATCC®25922 40 Turbid growth

Staphylococcus aureus ATCC®25923 78 Turbid growth

Control Media: Tryptone Soya Agar, Columbia Blood Agar Base enriched with 5% v/v horse blood, Columbia Blood Agar Base enriched with 5% v/v chocolate horse blood or Sabouraud Dextrose Agar, where appropriate

A satisfactory result is represented by visible growth from an inoculum of 10-100 colony-forming units (cfu).

For *Neisseria* species, a satisfactory result is represented by a positive diagnostic reaction, on subculture, from an inoculum of 10-100 cfu.

For Tube coagulase test, a satisfactory result is represented by the appropriate coagulase reaction from an inoculum of 1E+04 to 1E+06 cfu.

Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.04.28

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.