

# CERTIFICATE OF ANALYSIS

**PRODUCT** CT0020B  
E15 ERYTHROMYCIN

**LOT NUMBER** 6263280

**EXPIRY DATE** 2028.04.30

**DATE OF MANUFACTURE** 2025.05.01

## Delivery/Customer information

Date Printed

2025.05.15

Delivery No.

Customer

Customer Order Number

## Microbiological Performance

Assay (mcg/disc) 14.1

### Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	26	22 - 30
<i>Streptococcus pneumoniae</i> ATCC®49619	29	25 - 30
<i>Campylobacter jejuni</i> ATCC®33560	32	26 - 38

### Tested in accordance with current EUCAST methodology

<i>Staphylococcus aureus</i> ATCC®29213	24	23 - 29
<i>Streptococcus pneumoniae</i> ATCC®49619	30	26 - 32
<i>Haemophilus influenzae</i> ATCC®49766	16	10 - 16
<i>Campylobacter jejuni</i> ATCC®33560	28	27 - 35

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.15

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

**OXOID LIMITED**

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com