

# CERTIFICATE OF ANALYSIS

**PRODUCT** CT0052B  
SXT25 SULPHAMET/TRIMETHOP

**LOT NUMBER** 6282420

**EXPIRY DATE** 2028.06.26

**DATE OF MANUFACTURE** 2025.06.27

## Delivery/Customer information

Date Printed

2025.07.09

Delivery No.

Customer

Customer Order Number

## Microbiological Performance

Assay (mcg/disc) Sulphamethoxazole 32.440  
Assay (mcg/disc) Trimethoprim 1.810

### Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	28	24 - 32
<i>Escherichia coli</i> ATCC®25922	27	23 - 29
<i>Haemophilus influenzae</i> ATCC®49247	27	24 - 32
<i>Streptococcus pneumoniae</i> ATCC®49619	24	20 - 28

### Tested in accordance with current EUCAST methodology

<i>Escherichia coli</i> ATCC®25922	27	23 - 29
<i>Staphylococcus aureus</i> ATCC®29213	28	26 - 32
<i>Enterococcus faecalis</i> ATCC®29212	28	26 - 34
<i>Streptococcus pneumoniae</i> ATCC®49619	24	18 - 26
<i>Haemophilus influenzae</i> ATCC®49766	29	27 - 35

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.07.09

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.