

CERTIFICATE OF ANALYSIS

PRODUCT CT0052B
SXT25 SULPHAMET/TRIMETHOP

LOT NUMBER 6285668

EXPIRY DATE 2028.07.03

DATE OF MANUFACTURE 2025.07.04

Delivery/Customer information

Date Printed

2025.07.16

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc) Sulphamethoxazole 30.160
Assay (mcg/disc) Trimethoprim 1.740

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	29	24 - 32
<i>Escherichia coli</i> ATCC®25922	28	23 - 29
<i>Haemophilus influenzae</i> ATCC®49247	28	24 - 32
<i>Streptococcus pneumoniae</i> ATCC®49619	23	20 - 28

Tested in accordance with current EUCAST methodology

<i>Escherichia coli</i> ATCC®25922	28	23 - 29
<i>Staphylococcus aureus</i> ATCC®29213	27	26 - 32
<i>Enterococcus faecalis</i> ATCC®29212	31	26 - 34
<i>Streptococcus pneumoniae</i> ATCC®49619	25	18 - 26
<i>Haemophilus influenzae</i> ATCC®49766	31	27 - 35

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.07.16

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com