

CERTIFICATE OF ANALYSIS

PRODUCT	CT0052B SXT25 SULPHAMET/TRIMETHOP
LOT NUMBER	6326517
EXPIRY DATE	2028.10.16
DATE OF MANUFACTURE	2025.10.17

Delivery/Customer information

Date Printed
2025.10.30

Delivery No.

Customer
Customer Order Number

Microbiological Performance

Assay (mcg/disc) Sulphamethoxazole	31.020
Assay (mcg/disc) Trimethoprim	1.770

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	27	24 - 32
<i>Escherichia coli</i> ATCC®25922	28	23 - 29
<i>Haemophilus influenzae</i> ATCC®49247	28	24 - 32
<i>Streptococcus pneumoniae</i> ATCC®49619	24	20 - 28

Tested in accordance with current EUCAST methodology

<i>Escherichia coli</i> ATCC®25922	28	23 - 29
<i>Staphylococcus aureus</i> ATCC®29213	29	26 - 32
<i>Enterococcus faecalis</i> ATCC®29212	31	26 - 34
<i>Streptococcus pneumoniae</i> ATCC®49619	25	18 - 26
<i>Haemophilus influenzae</i> ATCC®49766	34	27 - 35

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.10.30

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.