



## **CERTIFICATE OF ANALYSIS**

PRODUCT	CT0425B CIP5 CIPROFLOXACIN	Delivery/Customer information Date Printed 2025.05.30
LOT NUMBER	6270462	Delivery No.
EXPIRY DATE	2028.05.19	Customer Customer Order Number
DATE OF MANUFACTURE	2025.05.20	

## **Microbiological Performance**

Assay (mcg/disc)

## Tested in accordance with current CLSI methodology

	Zone Size	Limits			
	( <b>mm</b> )	( <b>mm</b> )			
Staphylococcus aureus ATCC®25923	22	22 - 30			
Escherichia coli ATCC®25922	33	29 - 38			
Pseudomonas aeruginosa ATCC®27853	28	25 - 33			
Haemophilus influenzae ATCC®49247	36	34 - 42			
Neisseria gonorrhoeae ATCC®49226	50	48 - 58			
-					
Tested in accordance with current EUCAST methodology					
Escherichia coli ATCC®25922	33	29 - 37			
Pseudomonas aeruginosa ATCC®27853	28	25 - 33			
Staphylococcus aureus ATCC®29213	22	21 - 27			
Haemophilus influenzae ATCC®49766	34	32 - 40			
Streptococcus pneumoniae ATCC®49619	27	22 - 28			
Campylobacter jejuni ATCC®33560	39	34 - 42			

7.26

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.30

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection. NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.