



PRODUCTCT0425B
CIPROFLOXACINDelivery/Customer informationDate Printed
2025.06.18Delivery No.LOT NUMBER6275685Delivery No.EXPIRY DATE2028.06.09Customer
Customer Order NumberDATE OF
MANUFACTURE2025.06.10Customer
Customer Order Number

Microbiological Performance

Assay (mcg/disc)

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
Staphylococcus aureus ATCC®25923	22	22 - 30
Escherichia coli ATCC®25922	31	29 - 38
Pseudomonas aeruginosa ATCC®27853	30	25 - 33
Haemophilus influenzae ATCC®49247	34	34 - 42
Neisseria gonorrhoeae ATCC®49226	52	48 - 58
Tested in accordance with current EUCAST methodology		
Escherichia coli ATCC®25922	31	29 - 37
Pseudomonas aeruginosa ATCC®27853	30	25 - 33
Staphylococcus aureus ATCC®29213	21	21 - 27
Haemophilus influenzae ATCC®49766	36	32 - 40
Streptococcus pneumoniae ATCC®49619	24	22 - 28
Campylobacter jejuni ATCC®33560	42	34 - 42

5.91

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.18

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection. NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.