



PRODUCTCT0425B
CIPROFLOXACINDelivery/Customer information
Date Printed
2025.06.18
Delivery No.LOT NUMBER6280030Customer
Customer Order NumberEXPIRY DATE028.06.09Customer
Customer Order NumberDATE OF
MANUFACTURE025.06.10

Microbiological Performance

Assay (mcg/disc)

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
Staphylococcus aureus ATCC®25923	22	22 - 30
Escherichia coli ATCC®25922	31	29 - 38
Pseudomonas aeruginosa ATCC®27853	28	25 - 33
Haemophilus influenzae ATCC®49247	37	34 - 42
Neisseria gonorrhoeae ATCC®49226	52	48 - 58
Tested in accordance with current EUCAST methodology		
Escherichia coli ATCC®25922	31	29 - 37
Pseudomonas aeruginosa ATCC®27853	28	25 - 33
Staphylococcus aureus ATCC®29213	21	21 - 27
Haemophilus influenzae ATCC®49766	38	32 - 40
Streptococcus pneumoniae ATCC®49619	26	22 - 28
Campylobacter jejuni ATCC®33560	42	34 - 42

5.98

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.18

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection. NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.