

CERTIFICATE OF ANALYSIS

PRODUCT CT0425B

CIP5 CIPROFLOXACIN

LOT NUMBER 6281549

EXPIRY DATE 2028.06.23

DATE OF

MANUFACTURE 2025.06.24

Delivery/Customer information

Date Printed 2025.07.08

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc)

5.52

Tested in accordance with current CLSI methodology

Tested in determine with edition of the control of			
	Zone Size	Limits	
	(mm)	(mm)	
Staphylococcus aureus ATCC®25923	25	22 - 30	
Escherichia coli ATCC®25922	33	29 - 38	
Pseudomonas aeruginosa ATCC®27853	27	25 - 33	
Haemophilus influenzae ATCC®49247	35	34 - 42	
Neisseria gonorrhoeae ATCC®49226	55	48 - 58	
Tested in accordance with current EUCAST	methodology		
Escherichia coli ATCC®25922	33	29 - 37	
Pseudomonas aeruginosa ATCC®27853	27	25 - 33	
Staphylococcus aureus ATCC®29213	22	21 - 27	
Haemophilus influenzae ATCC®49766	32	32 - 40	
Streptococcus pneumoniae ATCC®49619	25	22 - 28	
Campylobacter jejuni ATCC®33560	41	34 - 42	

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.07.08

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.