

CERTIFICATE OF ANALYSIS

PRODUCT CT0446B

OFX5 OFLOXACIN

LOT NUMBER 6270464

EXPIRY DATE 2028.05.26

DATE OF

MANUFACTURE 2025.05.27

Microbiological Performance

Delivery/Customer information

Date Printed 2025.06.04

Delivery No.

Customer

Customer Order Number

Assay (mcg/disc)	7.3		
Tested in accordance with current CLSI met	hodology		
Staphylococcus aureus ATCC®25923	24	24 - 28	
Escherichia coli ATCC®25922	29	29 - 33	
Pseudomonas aeruginosa ATCC®27853	20	17 - 21	
Haemophilus influenzae ATCC®49247	35	31 - 40	
Streptococcus pneumoniae ATCC®49619	21	16 - 21	
Neisseria gonorrhoeae ATCC®49226	44	43 - 51	

Tested in accordance with current EUCAST methodology

Escherichia coli ATCC®25922

29	29 - 33
22	21 - 27
23	18 - 24
33	31 - 37
	22 23

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.04

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.