

CERTIFICATE OF ANALYSIS

PRODUCT CT0758B
FOT200 FOSFOMYCIN TROMETAMOL

LOT NUMBER 6280035

EXPIRY DATE 2027.06.18

DATE OF MANUFACTURE 2025.06.18

Delivery/Customer information

Date Printed

2025.06.27

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc) 229.99

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	28	25 - 33

Tested in accordance with current EUCAST methodology

PLEASE NOTE EUCAST have the following guidelines for reading the zone:

Ignore isolated colonies within the inhibition zone and read the outer zone edge

<i>Escherichia coli</i> ATCC®25922	31	26 - 34
------------------------------------	----	---------

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.27

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com