



CERTIFICATE OF ANALYSIS

PRODUCT	DD0001B OP OPTOCHIN (5X50 Discs)	Delivery/Customer information	
		Date Printed 2025.06.09	
		2023.06.09 Delivery No.	
LOT NUMBER	6272883		
EXPIRY DATE	2026.05.29	Customer	
	2020103125	Customer Order Number	
DATE OF MANUFACTURE	2025.05.29		

Microbiological Performance

Diagnostic characteristics			
	Zone Size	Limits	
	(mm)	(mm)	
Streptococcus pneumoniae ATCC®49619	20	16 - 27	
Streptococcus pyogenes ATCC®19615	0	0 - 0	

Quality control ranges are internally derived based on historical data.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.09

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.