

CERTIFICATE OF ANALYSIS

PRODUCT DD0004B
V FACTOR 5 x 50 DISCS

LOT NUMBER 6259601

EXPIRY DATE 2026.05.15

DATE OF MANUFACTURE 2025.05.15

Delivery/Customer information

Date Printed

2025.06.09

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Initial Inoculum

Test Result

Diagnostic characteristics

<i>Haemophilus influenzae</i> ATCC®19418	1E+04-1E+06	No growth
<i>Haemophilus influenzae</i> ATCC®49247	1E+04-1E+06	No growth
<i>Haemophilus parainfluenzae</i> ATCC®33392	1E+04-1E+06	Growth
<i>Haemophilus haemoglobinophilus</i> NCTC8540	1E+04-1E+06	No growth

Quality control ranges are internally derived based on historical data.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.09

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com