

Certificate of Analysis

Product Name: STAPHYTECT DRYSPOT KIT (120 TESTS)
Product Code: DR0100M
Lot Number: 6234926
Expiration Date: 30.04.2027
Date of Manufacture: 09.06.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74DR102A	Staphytest DrySpot Test Reagent	6184783	30.04.2027

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

Organism Challenge	OCC	Performance Criteria	ATCC	NCTC
S. Aureus	1009	Acceptable		
S. Aureus	1008	Acceptable		
S. Aureus	1011	Acceptable		
S. Aureus	198	Acceptable	25923	
S. Aureus	192	Acceptable	6538p	7447
S. Aureus	100	Acceptable		6571
S. Aureus	1027	Acceptable		
S. Aureus	1024	Acceptable		
S. Aureus	1133	Acceptable		
S. Aureus	2074	Acceptable		
S. Aureus	1150	Acceptable		
S. Aureus	934	Acceptable	43300	
S. Simulans	1020	Acceptable		
S. Epidermis	2038	Acceptable		
S. Haemolyticus	1016	Acceptable		
S. Epidermis	919	Acceptable	12228	
S. Epidermis	1630	Acceptable		
S. Hominis	2232	Acceptable		
S. Hominis	9503	Acceptable		

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.



Fay Curd
Quality Assurance Manager