

Certificate of Analysis

Product Name: EXTRACTION ENZYME 2 X 50 TESTS
Product Code: DR0593G
Lot Number: 6256584
Expiration Date: 31.07.2026
Date of Manufacture: 29.07.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74DR593G	New Strep Enzyme	6194098	31.07.2026

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

Organism Challenge	OCC	ATCC	NCTC	Performance Criteria
S. pyogenes	936			Acceptable
S. pyogenes	988			Acceptable
S. pyogenes	937			Acceptable
S. pyogenes	938			Acceptable
S. agalactiae	761	9993		Acceptable
S. agalactiae	765	11360		Acceptable
S. agalactiae	869			Acceptable
Streptococcus sp.	870			Acceptable
S. equisimilis	308	35666	8543	Acceptable
S. equisimilis	950	5371		Acceptable
S. equisimilis	951			Acceptable
S. equisimilis	953			Acceptable
S. faecalis	640	29212		Acceptable
S. faecalis	957			Acceptable
S. faecalis	959			Acceptable
S. faecalis	965			Acceptable
Streptococcus sp.	970			Acceptable
S. milleri	971			Acceptable
S. milleri	972			Acceptable
S. milleri	973			Acceptable
S. canis	975			Acceptable
S. canis	976			Acceptable
S. canis	977			Acceptable
S. canis	978			Acceptable

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.



Fay Curd
Quality Assurance Manager