

Certificate of Analysis

Product Name: STAPHYLASE TEST 1 X 100 TESTS

Product Code: DR0595A Lot Number: 6205107 Expiration Date: 31.10.2025 Date of Manufacture: 15.04.2025

*Date format is DD.MM.YYYY

 Item
 Description
 Lot/Serial
 Expire

 W74DR596A
 Staphylase Test Reagent
 6186666
 31.10.2025

 W74DR597A
 Staphylase Control Reagent
 6186667
 31.10.2025

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

Organism Challenge	OCC	ATCC NCTC	Performance Criteria
S. Aureus	1009		Acceptable
S. Aureus	198	25923	Acceptable
S. Aureus	192	6539p 7447	Acceptable
S. Aureus	1012		Acceptable
S. hominis	1022		Acceptable
S. epidermis	691	14990 11047	Acceptable

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.

Fay Curd

Quality Assurance Manager