

Certificate of Analysis

Product Name: STAPHYLASE TEST 1 X 100 TESTS
Product Code: DR0595A
Lot Number: 6233542
Expiration Date: 21.12.2025
Date of Manufacture: 21.05.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74DR596A	Staphylase Test Reagent	6205108	21.12.2025
W74DR597A	Staphylase Control Reagent	6205109	21.12.2025

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

Organism Challenge	OCC	ATCC	NCTC	Performance Criteria
S. Aureus	1009			Acceptable
S. Aureus	198	25923		Acceptable
S. Aureus	192	6539p	7447	Acceptable
S. Aureus	1012			Acceptable
S. hominis	1022			Acceptable
S. epidermis	691	14990	11047	Acceptable

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.



Fay Curd
Quality Assurance Manager