

Certificate of Analysis

Product Name:PathoDxtra Strep Group A LatexProduct Code:DR0701GLot Number:6186228Expiration Date:31.05.2026Date of Manufacture:08.05.2025

*Date format is DD.MM.YYYY

ltem W74PDX70A	Description Strep A Group Latex	Lot/Serial 6135357	Expire 31.05.2026
Test results carried out according to the Quality Assurance Product Release and Test Specification			
Test	Anticipated Result		Actual Result
Latex Efficacy/ Specificity using Extraction Reagents	The Latex must show obvious agglutination with its homologous extract		Conforms
	The Latex must not show agglutination with the heterologous extracts but a trace reaction is acceptable		Conforms
Reactivity of Latex with Positive Control	The Latex must show strong positive agglutination with the positive control.		Conforms

The

Fay Curd Quality Assurance Manager