

## Certificate of Analysis

**Product Name:** PathoDextra Strep Group B Latex  
**Product Code:** DR0702G  
**Lot Number:** 6196184  
**Expiration Date:** 31.07.2026  
**Date of Manufacture:** 02.06.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74PDX71B	Strep B Group Latex	6169136	31.07.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Latex Efficacy/ Specificity using Extraction Reagents	The Latex must show obvious agglutination with its homologous extract	Conforms
	The Latex must not show agglutination with the heterologous extracts but a trace reaction is acceptable	Conforms
Reactivity of Latex with Positive Control	The Latex must show strong positive agglutination with the positive control.	Conforms



Fay Curd  
Quality Assurance Manager