

## **Certificate of Analysis**

Product Name:PathoDxtra Strep Group FProduct Code:DR0705GLot Number:6215328Expiration Date:31.07.2026Date of Manufacture:23.04.2025

\*Date format is DD.MM.YYYY

<b>ltem</b> W74PDX74F	Description Strep F Group Latex	Lot/Serial 6186238	<b>Expire</b> 31.07.2026
Test results carried out according to the Quality Assurance Product Release and Test Specification			
Test	Anticipated Result		Actual Result
Latex Efficacy/ Specificity using Extraction Reagents	The Latex must show obvious agglutination with its homologous extract		Conforms
	The Latex must not show agglutination with the heterologous extracts but a trace reaction is acceptable		Conforms
Reactivity of Latex with Positive Control	The Latex must show strong positive agglutination with the positive control.		Conforms

The

Fay Curd Quality Assurance Manager