

Certificate of Analysis

Product Name: PathoDxtra Strep Group F
Product Code: DR0705G
Lot Number: 6215329
Expiration Date: 31.07.2026
Date of Manufacture: 15.07.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74PDX74F	Strep F Group Latex	6186238	31.07.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Latex Efficacy/ Specificity using Extraction Reagents	The Latex must show obvious agglutination with its homologous extract	Conforms
	The Latex must not show agglutination with the heterologous extracts but a trace reaction is acceptable	Conforms
Reactivity of Latex with Positive Control	The Latex must show strong positive agglutination with the positive control.	Conforms



Fay Curd
Quality Assurance Manager