

CERTIFICATE OF ANALYSIS

PRODUCT CM0041B

SABOURAUD DEXTROSE AGAR 500g

LOT NUMBER

6252207

EXPIRY DATE

2030.04.30

DATE OF

MANUFACTURE 2025.04.01

Delivery/Customer information

Date Printed 2025.04.16

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results Straw powder		Specification Straw powder	
Appearance				
Colour on reconstitution	Straw 2		Straw 1-2	
pH (25°C)	5.6		5.4 - 5.8	
Clarity	Clear		Clear	
Microbiological Performance	Control	Test	Recovery	Test
	cfu	cfu	%	Description
Aerobic incubation at 20-25°C for up to 5 days				
Saccharomyces carlsbergensis ATCC®2700	48	48	100	Cream, domed cols
Candida albicans ATCC®10231	40	46	115	Cream, domed cols
Aspergillus brasiliensis ATCC®16404	92	85	92	White mycelia, black spores
Testing has been performed in accordance with I	SO11133:2014			
Aerobic incubation at $25 \pm 2^{\circ}$ C for 5 days				
S. cerevisiae ATCC®9763 WDCM00058	72	81	113	Cream, domed cols
A. brasiliensis ATCC®16404 WDCM00053	92	85	92	White mycelia, black spores

Control Medium: Sabouraud Dextrose Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium. Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.04.16

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.