



CERTIFICATE OF ANALYSIS

PRODUCT CM0129T

TRYPTONE SOYA BROTH 5kg

LOT NUMBER 6246650

EXPIRY DATE 2030.04.30

DATE OF

MANUFACTURE 2025.04.07

Delivery/Customer information

Date Printed 2025.04.22

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification	
Appearance	Straw powder	Straw powder	
Colour on reconstitution	Straw 3	Straw 2-3	
pH (25°C)	7.1	7.1 - 7.5	
Clarity	Clear	Clear	
Thermophiles & mesophiles	Absent	Absent	
Microbiological Performance	Initial Inoculum	Test Result	
Aerobic incubation at 30-35°C for 18-24 hours			
Streptococcus pyogenes ATCC®19615	31	Turbid growth	
Tested in accordance with current USP/EP/BP/JI	P		
Aerobic incubation at 30-35°C for 24 hours			
Escherichia coli ATCC®8739	70	Turbid growth	
Staphylococcus aureus ATCC®6538	54	Turbid growth	
Pseudomonas aeruginosa ATCC®9027	80	Turbid growth	
Salmonella typhimurium ATCC®14028	63	Turbid growth	
Salmonella abony NCTC6017	85	Turbid growth	
Aerobic incubation at 30-35°C for up to 3 days			
Bacillus subtilis ATCC®6633	50	Flocculent/surface growth	
Kocuria rhizophila ATCC®9341	73	Turbid growth	
Aerobic incubation at 20-25°C for 48 hours			
Bacillus subtilis ATCC®6633	50	Flocculent/surface growth	
Candida albicans ATCC®10231	55	Flocculent/surface growth	
Aerobic incubation at 20-25°C for 5 days			
Aspergillus brasiliensis ATCC®16404	46	White mycelia, no spores	





CERTIFICATE OF ANALYSIS

PRODUCT CM0129T

TRYPTONE SOYA BROTH 5kg

LOT NUMBER 6246650

Tested in accordance with current CLSI M22 A Aerobic incubation at 35°C for 18-24 hours

Escherichia coli ATCC®25922 1E+04-1E+06 Turbid growth Staphylococcus aureus ATCC®25923 1E+04-1E+06 Turbid growth

Aerobic incubation at 35°C for up to 5 days

Streptococcus pneumoniae ATCC®6305 1E+04-1E+06 Turbid growth

Anaerobic incubation at 35°C for up to 5 days

Bacteroides fragilis ATCC®25285 1E+04-1E+06 Turbid growth

Control Media: Tryptone Soya Agar, Columbia Blood Agar Base enriched with 5% v/v horse blood or Sabouraud Dextrose Agar where appropriate

A satisfactory result is represented by visible growth from an inoculum of 10-100 colony-forming units (cfu). Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.04.22

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.