



CERTIFICATE OF ANALYSIS

PRODUCT CM0129T

TRYPTONE SOYA BROTH 5kg

LOT NUMBER 6271346

EXPIRY DATE 2030.05.31

DATE OF

MANUFACTURE 2025.05.14

Delivery/Customer information

Date Printed 2025.05.31

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 2	Straw 2-3
pH (25°C)	7.2	7.1 - 7.5
Clarity	Clear	Clear
Thermophiles & mesophiles	Absent	Absent
Microbiological Performance	Initial Inoculum	Test Result
Aerobic incubation at 30-35°C for 18-24 hours		
Streptococcus pyogenes ATCC®19615	45	Turbid growth
Tested in accordance with current USP/EP/BP/J	P	
Aerobic incubation at 30-35°C for 24 hours		
Escherichia coli ATCC®8739	61	Turbid growth
Staphylococcus aureus ATCC®6538	42	Turbid growth
Pseudomonas aeruginosa ATCC®9027	83	Turbid growth
Salmonella typhimurium ATCC®14028	58	Turbid growth
Salmonella abony NCTC6017	75	Turbid growth
Aerobic incubation at 30-35°C for up to 3 days		
Bacillus subtilis ATCC®6633	95	Flocculent/surface growth
Kocuria rhizophila ATCC®9341	75	Turbid growth
Aerobic incubation at 20-25°C for 48 hours		
Bacillus subtilis ATCC®6633	95	Flocculent/surface growth
Candida albicans ATCC®10231	59	Flocculent/surface growth
Aerobic incubation at 20-25°C for 5 days		
Aspergillus brasiliensis ATCC®16404	17	White mycelia, no spores





CERTIFICATE OF ANALYSIS

PRODUCT CM0129T

TRYPTONE SOYA BROTH 5kg

LOT NUMBER 6271346

Tested in accordance with current CLSI M22 A Aerobic incubation at 35°C for 18-24 hours

Escherichia coli ATCC®25922 1E+04-1E+06 Turbid growth Staphylococcus aureus ATCC®25923 1E+04-1E+06 Turbid growth

Aerobic incubation at 35°C for up to 5 days

Streptococcus pneumoniae ATCC®6305 1E+04-1E+06 Turbid growth

Anaerobic incubation at 35°C for up to 5 days

Bacteroides fragilis ATCC®25285 1E+04-1E+06 Turbid growth

Control Media: Tryptone Soya Agar, Columbia Blood Agar Base enriched with 5% v/v horse blood or Sabouraud Dextrose Agar where appropriate

A satisfactory result is represented by visible growth from an inoculum of 10-100 colony-forming units (cfu). Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.30

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.