



## **CERTIFICATE OF ANALYSIS**

PRODUCT CM0139B

POTATO DEXTROSE AGAR 500g

LOT NUMBER

6243709

**EXPIRY DATE** 

2030.03.31

**DATE OF** 

**MANUFACTURE** 2025.03.12

Delivery/Customer information

Date Printed 2025.05.07

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results		Specifica	tion
Appearance	Off white powder		Off white powder	
Colour on reconstitution	Straw 1		Straw 1	
pH (25°C)	5.7		5.4 - 5.8	
pH of acidified medium (25°C)	3.6		3.4 - 3.6	
Clarity	Clear		Clear	
SR0021 per litre to lower pH to 3.5 (ml)	20		10 - 100	
Microbiological Performance	Control	Test	Recovery	Test
	cfu	cfu	%	Description
Tested using acidified medium				
Aerobic incubation at 25°C for 5 days				
Candida krusei ATCC®6258	59	58	98	Cream cols
Candida albicans ATCC®10231	35	40	114	Cream cols
Bacillus subtilis ATCC®6633	>1E+04	0	-	No growth
Escherichia coli ATCC®8739	>1E+04	0	_	No growth
Staphylococcus aureus ATCC®6538	>1E+04	0	-	No growth
Inoculation using stab technique				
Aspergillus fumigatus ATCC®9197	>1E+04	_	-	Cream mycelia, blue/green spores
Aspergillus brasiliensis ATCC®16404	>1E+04	_	-	White mycelia, black spores
Fusarium culmorum FMH-2407W	>1E+04	_	-	White mycelia, orange/red spores
Fusarium oxysporum FYH-2386W	>1E+04	-	-	White mycelia, pink/red spores
Tested in accordance with USP/EP/BP/JP				
Aerobic incubation at 25°C for 5 to 7 days Aspergillus brasiliensis ATCC®16404	47	46	98	White mycelia, black spores

Control Medium: Potato Dextrose Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu). Negative strains are inhibited from an inoculum of 1E+04 to 1E+06 cfu. For mould species, a satisfactory result is represented by typical colonial morphology as described in the product specification. Refer to product specification for full details.





## **CERTIFICATE OF ANALYSIS**

PRODUCT CM0139B

POTATO DEXTROSE AGAR 500g

**LOT NUMBER** 6243709

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.07

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.