

CERTIFICATE OF ANALYSIS

PRODUCT CM0151B
REINFORCED CLOSTRIDIAL AGAR 500g

LOT NUMBER 6257698

EXPIRY DATE 2030.04.30

DATE OF MANUFACTURE 2025.04.15

Delivery/Customer information

Date Printed

2025.05.06

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 2	Straw 2-3
pH (25°C)	6.7	6.6 - 7.0
Clarity	Clear	Clear
Thermophiles	Absent	Absent

Microbiological Performance	Initial Inoculum	Test Result
Aerobic incubation at 37°C for 24-48 hours		
<i>Clostridium perfringens</i> ATCC®13124	91	Growth & gas
<i>Clostridium sporogenes</i> ATCC®19404	81	Growth & gas
<i>Clostridium tetani</i> ATCC®9441	39	Growth & gas
<i>Clostridium sphenoides</i> ATCC®19403	57	Growth & gas

Control Medium: Columbia Blood Agar Base enriched with 5% v/v horse blood

A satisfactory result is represented by visible growth and gas from an inoculum of 10-100 colony-forming units (cfu).
Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.06

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.